

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
 Data File : VX026517.D
 Acq On : 19 Jan 2022 14:24
 Operator : JC/MD
 Sample : N1187-08 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31338

Quant Time: Jan 20 05:05:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	138264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	233919	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200922	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84604	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94738	49.050	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.100%
35) Dibromofluoromethane	5.391	113	74144	49.083	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.160%
50) Toluene-d8	8.653	98	269775	49.559	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.120%
62) 4-Bromofluorobenzene	11.085	95	94834	49.042	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
Data File : VX026517.D
Acq On : 19 Jan 2022 14:24
Operator : JC/MD
Sample : N1187-08 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31338

Quant Time: Jan 20 05:05:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

