

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
 Data File : VX026519.D
 Acq On : 19 Jan 2022 15:10
 Operator : JC/MD
 Sample : N1187-06 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31282

Quant Time: Jan 20 05:06:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	231119	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196768	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83094	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93625	49.255	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.520%
35) Dibromofluoromethane	5.391	113	73521	49.260	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.520%
50) Toluene-d8	8.653	98	265953	49.449	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.900%
62) 4-Bromofluorobenzene	11.079	95	94244	49.327	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.660%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	173304	178.786	ug/l	99
18) Methyl Acetate	2.709	43	942	0.489	ug/l #	85
25) 2-Butanone	4.574	43	2302	1.826	ug/l	99
95) Naphthalene	13.780	128	1933	0.327	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
Data File : VX026519.D
Acq On : 19 Jan 2022 15:10
Operator : JC/MD
Sample : N1187-06 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31282

Quant Time: Jan 20 05:06:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

