

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX011922\
 Data File : VX026507.D
 Acq On : 19 Jan 2022 10:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 20 05:01:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	45.733	8.5	92	0.00
3 P	Chloromethane	50.000	48.778	2.4	101	0.00
4 C	Vinyl Chloride	50.000	48.516	3.0#	96	0.00
5 T	Bromomethane	50.000	52.905	-5.8	108	0.00
6 T	Chloroethane	50.000	45.037	9.9	91	0.00
7 T	Trichlorofluoromethane	50.000	47.265	5.5	92	0.00
8 T	Diethyl Ether	50.000	48.826	2.3	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.266	1.5	96	0.00
10 T	Methyl Iodide	50.000	56.911	-13.8	120	0.00
11 T	Tert butyl alcohol	250.000	230.757	7.7	88	0.00
12 CM	1,1-Dichloroethene	50.000	48.447	3.1#	95	0.00
13 T	Acrolein	250.000	262.102	-4.8	111	0.00
14 T	Allyl chloride	50.000	48.813	2.4	94	0.00
15 T	Acrylonitrile	250.000	237.180	5.1	94	0.00
16 T	Acetone	250.000	224.905	10.0	88	0.00
17 T	Carbon Disulfide	50.000	46.348	7.3	94	0.00
18 T	Methyl Acetate	50.000	47.431	5.1	96	0.00
19 T	Methyl tert-butyl Ether	50.000	47.249	5.5	93	0.00
20 T	Methylene Chloride	50.000	47.523	5.0	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.100	3.8	94	0.00
22 T	Diisopropyl ether	50.000	48.748	2.5	95	0.00
23 T	Vinyl Acetate	250.000	245.593	1.8	93	0.00
24 P	1,1-Dichloroethane	50.000	48.730	2.5	95	0.00
25 T	2-Butanone	250.000	239.744	4.1	92	0.00
26 T	2,2-Dichloropropane	50.000	49.162	1.7	92	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.094	1.8	95	0.00
28 T	Bromochloromethane	50.000	52.326	-4.7	101	0.00
29 T	Tetrahydrofuran	250.000	231.484	7.4	92	0.00
30 C	Chloroform	50.000	49.308	1.4#	97	0.00
31 T	Cyclohexane	50.000	47.398	5.2	95	0.00
32 T	1,1,1-Trichloroethane	50.000	49.035	1.9	93	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.403	5.2	93	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	100	0.00
35 S	Dibromofluoromethane	50.000	49.198	1.6	94	0.00
36 T	1,1-Dichloropropene	50.000	49.153	1.7	95	0.00
37 T	Ethyl Acetate	50.000	48.597	2.8	93	0.00
38 T	Carbon Tetrachloride	50.000	50.617	-1.2	95	0.00
39 T	Methylcyclohexane	50.000	50.940	-1.9	98	0.00
40 TM	Benzene	50.000	49.968	0.1	97	0.00
41 T	Methacrylonitrile	50.000	49.334	1.3	95	0.00
42 TM	1,2-Dichloroethane	50.000	48.529	2.9	94	0.00
43 T	Isopropyl Acetate	50.000	48.855	2.3	92	0.00
44 TM	Trichloroethene	50.000	51.446	-2.9	99	0.00
45 C	1,2-Dichloropropane	50.000	50.637	-1.3#	97	0.00
46 T	Dibromomethane	50.000	50.115	-0.2	96	0.00
47 T	Bromodichloromethane	50.000	51.123	-2.2	96	0.00
48 T	Methyl methacrylate	50.000	48.865	2.3	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
 Data File : VX026507.D
 Acq On : 19 Jan 2022 10:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 20 05:01:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	888.170	11.2	84	0.00
50 S	Toluene-d8	50.000	50.152	-0.3	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	244.235	2.3	93	0.00
52 CM	Toluene	50.000	50.786	-1.6#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	47.815	4.4	97	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.474	-6.9	97	0.00
55 T	1,1,2-Trichloroethane	50.000	50.984	-2.0	98	0.00
56 T	Ethyl methacrylate	50.000	51.004	-2.0	94	0.00
57 T	1,3-Dichloropropane	50.000	50.825	-1.7	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	253.071	-1.2	96	0.00
59 T	2-Hexanone	250.000	255.253	-2.1	96	0.00
60 T	Dibromochloromethane	50.000	52.818	-5.6	97	0.00
61 T	1,2-Dibromoethane	50.000	51.056	-2.1	98	0.00
62 S	4-Bromofluorobenzene	50.000	50.560	-1.1	97	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	100	0.00
64 T	Tetrachloroethene	50.000	52.442	-4.9	103	0.00
65 PM	Chlorobenzene	50.000	49.947	0.1	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.768	-3.5	99	0.00
67 C	Ethyl Benzene	50.000	51.165	-2.3#	98	0.00
68 T	m/p-Xylenes	100.000	103.241	-3.2	99	0.00
69 T	o-Xylene	50.000	52.135	-4.3	100	0.00
70 T	Styrene	50.000	52.162	-4.3	97	0.00
71 P	Bromoform	50.000	48.308	3.4	100	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	101	0.00
73 T	Isopropylbenzene	50.000	49.610	0.8	98	0.00
74 T	N-amyl acetate	50.000	49.270	1.5	92	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.812	4.4	95	0.00
76 T	1,2,3-Trichloropropane	50.000	48.902	2.2	95	0.00
77 T	Bromobenzene	50.000	48.880	2.2	97	0.00
78 T	n-propylbenzene	50.000	50.633	-1.3	99	0.00
79 T	2-Chlorotoluene	50.000	48.917	2.2	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.128	-0.3	99	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.326	3.3	102	0.00
82 T	4-Chlorotoluene	50.000	49.174	1.7	98	0.00
83 T	tert-Butylbenzene	50.000	50.436	-0.9	98	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.007	-0.0	98	0.00
85 T	sec-Butylbenzene	50.000	51.259	-2.5	99	0.00
86 T	p-Isopropyltoluene	50.000	52.081	-4.2	100	0.00
87 T	1,3-Dichlorobenzene	50.000	49.751	0.5	101	0.00
88 T	1,4-Dichlorobenzene	50.000	49.149	1.7	98	0.00
89 T	n-Butylbenzene	50.000	52.710	-5.4	101	0.00
90 T	Hexachloroethane	50.000	48.799	2.4	101	0.00
91 T	1,2-Dichlorobenzene	50.000	50.157	-0.3	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.868	2.3	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.591	-3.2	100	0.00
94 T	Hexachlorobutadiene	50.000	51.453	-2.9	102	0.00
95 T	Naphthalene	50.000	50.719	-1.4	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011922\
Data File : VX026507.D
Acq On : 19 Jan 2022 10:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 20 05:01:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.318	-2.6	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6