

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011923\
 Data File : VX033792.D
 Acq On : 19 Jan 2023 11:42
 Operator : JC/MD
 Sample : PB150330TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150330TB

Quant Time: Jan 20 05:19:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	170312	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	245182	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213746	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92363	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82850	46.202	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.400%
35) Dibromofluoromethane	5.385	113	74056	48.643	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.280%
50) Toluene-d8	8.647	98	280236	50.519	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.040%
62) 4-Bromofluorobenzene	11.079	95	93248	45.276	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.560%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	24949	27.245	ug/l	95
17) Carbon Disulfide	2.508	76	868	2.061	ug/l #	83
37) Ethyl Acetate	4.715	43	27396	10.674	ug/l	98
43) Isopropyl Acetate	6.336	43	111913	31.904	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011923\
Data File : VX033792.D
Acq On : 19 Jan 2023 11:42
Operator : JC/MD
Sample : PB150330TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150330TB

Quant Time: Jan 20 05:19:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

