

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011923\
 Data File : VX033796.D
 Acq On : 19 Jan 2023 13:15
 Operator : JC/MD
 Sample : PB150330ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150330ZHE#04

Quant Time: Jan 20 05:20:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	173019	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	250162	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214590	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95441	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83429	45.797	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.600%
35) Dibromofluoromethane	5.385	113	75618	48.680	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.360%
50) Toluene-d8	8.647	98	283652	50.117	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.240%
62) 4-Bromofluorobenzene	11.079	95	94199	44.828	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.660%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	24489	25.884	ug/l	91
20) Methylene Chloride	2.788	84	12251	5.505	ug/l	94
37) Ethyl Acetate	4.715	43	22726	8.111	ug/l	100
43) Isopropyl Acetate	6.336	43	99456	27.789	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011923\
Data File : VX033796.D
Acq On : 19 Jan 2023 13:15
Operator : JC/MD
Sample : PB150330ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150330ZHE#04

Quant Time: Jan 20 05:20:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

