

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011923\
 Data File : VX033798.D
 Acq On : 19 Jan 2023 14:01
 Operator : JC/MD
 Sample : PB150330ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150330ZHE#06

Quant Time: Jan 20 05:21:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	171647	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	248818	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99563	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83474	46.188	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.380%
35) Dibromofluoromethane	5.385	113	75137	48.632	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.260%
50) Toluene-d8	8.647	98	282250	50.139	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.280%
62) 4-Bromofluorobenzene	11.079	95	94299	45.118	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.240%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	25731	28.185	ug/l	91
20) Methylene Chloride	2.788	84	11988	5.394	ug/l	95
37) Ethyl Acetate	4.715	43	21699	7.665	ug/l	99
43) Isopropyl Acetate	6.336	43	99636	27.989	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011923\
Data File : VX033798.D
Acq On : 19 Jan 2023 14:01
Operator : JC/MD
Sample : PB150330ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150330ZHE#06

Quant Time: Jan 20 05:21:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

