

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011923\
 Data File : VX033803.D
 Acq On : 19 Jan 2023 15:57
 Operator : JC/MD
 Sample : PB150330ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150330ZHE#03

Quant Time: Jan 20 05:22:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	172557	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	254600	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222309	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85961	47.313	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.620%
35) Dibromofluoromethane	5.385	113	75469	47.737	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.480%
50) Toluene-d8	8.647	98	287337	49.883	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.079	95	100648	47.062	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.120%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	25601	27.760	ug/l	96
20) Methylene Chloride	2.788	84	12030	5.379	ug/l	94
37) Ethyl Acetate	4.715	43	21880	7.510	ug/l	98
43) Isopropyl Acetate	6.336	43	100248	27.522	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011923\
Data File : VX033803.D
Acq On : 19 Jan 2023 15:57
Operator : JC/MD
Sample : PB150330ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150330ZHE#03

Quant Time: Jan 20 05:22:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

