

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX011923\  
 Data File : VX033813.D  
 Acq On : 19 Jan 2023 19:49  
 Operator : JC/MD  
 Sample : 01220-05  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 28 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TP-G

Quant Time: Jan 20 05:24:36 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X011123W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jan 11 14:48:42 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 175692   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 260011   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 230461   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 112849   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 87088    | 47.078   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =      | 94.160%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 78543    | 48.648   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 97.300%  |
| 50) Toluene-d8              | 8.647  | 98    | 295248   | 50.190   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =      | 100.380% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 103307   | 47.300   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 94.600%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 2.392  | 43    | 24400    | 25.153   | ug/l   | 94       |
| 37) Ethyl Acetate           | 4.715  | 43    | 24406    | 8.482    | ug/l   | 99       |
| 43) Isopropyl Acetate       | 6.336  | 43    | 107694   | 28.951   | ug/l   | 98       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX011923\  
Data File : VX033813.D  
Acq On : 19 Jan 2023 19:49  
Operator : JC/MD  
Sample : 01220-05  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 28 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TP-G

Quant Time: Jan 20 05:24:36 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X011123W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jan 11 14:48:42 2023  
Response via : Initial Calibration

