

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039825.D
 Acq On : 19 Jan 2024 09:04
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050

Quant Time: Jan 20 00:19:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	96940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	170299	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	157761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	77864	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	75787	46.670	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.340%
35) Dibromofluoromethane	5.379	113	56282	49.067	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.140%
50) Toluene-d8	8.647	98	210366	48.627	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.260%
62) 4-Bromofluorobenzene	11.079	95	85731	50.677	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.360%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	53380	42.560	ug/l	97
3) Chloromethane	1.295	50	58621	40.925	ug/l	99
4) Vinyl Chloride	1.374	62	59798	44.236	ug/l	98
5) Bromomethane	1.599	94	38345	49.714	ug/l	98
6) Chloroethane	1.685	64	36636	47.674	ug/l	97
7) Trichlorofluoromethane	1.886	101	96379	49.424	ug/l	99
8) Diethyl Ether	2.130	74	34285	47.576	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.331	101	58099	48.057	ug/l	98
10) Methyl Iodide	2.453	142	65198	51.120	ug/l	95
11) Tert butyl alcohol	2.953	59	79525	188.130	ug/l #	92
12) 1,1-Dichloroethene	2.319	96	54881	47.038	ug/l	91
13) Acrolein	2.233	56	56528	215.297	ug/l	100
14) Allyl chloride	2.660	41	104466	45.492	ug/l	99
15) Acrylonitrile	3.056	53	185646	212.610	ug/l	99
16) Acetone	2.374	43	215109	235.539	ug/l	94
17) Carbon Disulfide	2.508	76	155651	46.415	ug/l	99
18) Methyl Acetate	2.697	43	86593	45.321	ug/l	100
19) Methyl tert-butyl Ether	3.105	73	192900	46.570	ug/l	99
20) Methylene Chloride	2.788	84	62568	45.293	ug/l	97
21) trans-1,2-Dichloroethene	3.087	96	59106	48.139	ug/l	98
22) Diisopropyl ether	3.751	45	203118	46.350	ug/l	95
23) Vinyl Acetate	3.715	43	956985	234.925	ug/l	100
24) 1,1-Dichloroethane	3.605	63	112992	46.868	ug/l	99
25) 2-Butanone	4.544	43	281101	214.185	ug/l	99
26) 2,2-Dichloropropane	4.471	77	93953	48.556	ug/l	98
27) cis-1,2-Dichloroethene	4.483	96	68505	48.646	ug/l	97
28) Bromochloromethane	4.891	49	52586	48.119	ug/l	98
29) Tetrahydrofuran	4.989	42	174385	206.438	ug/l	99
30) Chloroform	5.087	83	117206	49.247	ug/l	98
31) Cyclohexane	5.464	56	99606	46.477	ug/l	100
32) 1,1,1-Trichloroethane	5.379	97	100910	49.526	ug/l	99
36) 1,1-Dichloropropene	5.690	75	87469	48.991	ug/l	99
37) Ethyl Acetate	4.702	43	101408	42.575	ug/l	99
38) Carbon Tetrachloride	5.672	117	84836	51.880	ug/l	97
39) Methylcyclohexane	7.373	83	104714	48.842	ug/l	97
40) Benzene	6.031	78	243380	48.140	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039825.D
 Acq On : 19 Jan 2024 09:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jan 20 00:19:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	54066	46.418	ug/l	95
42) 1,2-Dichloroethane	6.080	62	96637	50.278	ug/l	100
43) Isopropyl Acetate	6.330	43	165472	47.358	ug/l	99
44) Trichloroethene	7.117	130	62903	50.768	ug/l	98
45) 1,2-Dichloropropane	7.421	63	64791	47.773	ug/l	100
46) Dibromomethane	7.574	93	47603	50.818	ug/l	97
47) Bromodichloromethane	7.818	83	90779	52.325	ug/l	94
48) Methyl methacrylate	7.684	41	80712	46.995	ug/l	93
49) 1,4-Dioxane	7.653	88	27024	813.181	ug/l #	92
51) 4-Methyl-2-Pentanone	8.568	43	547512	227.521	ug/l	98
52) Toluene	8.714	92	153270	50.757	ug/l	94
53) t-1,3-Dichloropropene	8.976	75	99415	54.003	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	105529	52.133	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	63803	49.627	ug/l	96
56) Ethyl methacrylate	9.110	69	103764	49.867	ug/l	96
57) 1,3-Dichloropropane	9.305	76	107053	49.135	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	231486	235.359	ug/l	96
59) 2-Hexanone	9.427	43	445716	234.026	ug/l	98
60) Dibromochloromethane	9.519	129	67690	55.829	ug/l	100
61) 1,2-Dibromoethane	9.604	107	68777	51.659	ug/l	100
64) Tetrachloroethene	9.269	164	52470	51.144	ug/l	96
65) Chlorobenzene	10.073	112	163045	49.621	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	59483	53.165	ug/l	98
67) Ethyl Benzene	10.189	91	308638	50.594	ug/l	98
68) m/p-Xylenes	10.299	106	226104	100.782	ug/l	94
69) o-Xylene	10.640	106	111008	49.964	ug/l	94
70) Styrene	10.653	104	186233	51.313	ug/l	98
71) Bromoform	10.799	173	46954	55.890	ug/l #	99
73) Isopropylbenzene	10.957	105	302540	48.803	ug/l	100
74) N-amyl acetate	10.842	43	153938	47.638	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.207	83	108269	45.737	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	119980	59.434	ug/l	91
77) Bromobenzene	11.195	156	68001	49.266	ug/l	98
78) n-propylbenzene	11.299	91	374974	49.975	ug/l	97
79) 2-Chlorotoluene	11.360	91	219920	47.423	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	261094	49.589	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.018	75	31978	47.556	ug/l	96
82) 4-Chlorotoluene	11.451	91	259757	49.382	ug/l	98
83) tert-Butylbenzene	11.713	119	250675	49.869	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	263851	50.230	ug/l	100
85) sec-Butylbenzene	11.890	105	335269	51.381	ug/l	99
86) p-Isopropyltoluene	12.006	119	270673	51.902	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	133053	50.287	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	133214	48.748	ug/l	99
89) n-Butylbenzene	12.329	91	273457	54.290	ug/l	95
90) Hexachloroethane	12.536	117	47146	53.587	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	127792	48.855	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	27199	46.889	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	83520	52.248	ug/l	98
94) Hexachlorobutadiene	13.725	225	36904	55.537	ug/l	97
95) Naphthalene	13.774	128	283144	48.115	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	80135	49.671	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039825.D
Acq On : 19 Jan 2024 09:04
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Quant Time: Jan 20 00:19:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 20 00:19:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:168 Resp: 96940
Ion Ratio Lower Upper
168 100
99 60.4 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 42.560 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 85 Resp: 53380
Ion Ratio Lower Upper
85 100
87 31.4 16.7 50.1









#3
Chloromethane
Concen: 40.925 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 58621
Ion Ratio Lower Upper
50 100
52 32.7 25.5 38.3



#4
Vinyl Chloride
Concen: 44.236 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 62 Resp: 59798
Ion Ratio Lower Upper
62 100
64 30.8 25.4 38.0









#5
Bromomethane
Concen: 49.714 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 38345
Ion Ratio Lower Upper
94 100
96 94.9 74.6 111.8



#6
Chloroethane
Concen: 47.674 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 64 Resp: 36636
Ion Ratio Lower Upper
64 100
66 33.2 25.1 37.7









#7
Trichlorofluoromethane
Concen: 49.424 ug/l
RT: 1.886 min Scan# 131
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 96379
Ion Ratio Lower Upper
101 100
103 63.6 51.4 77.2



#8
Diethyl Ether
Concen: 47.576 ug/l
RT: 2.130 min Scan# 171
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 74 Resp: 34285
Ion Ratio Lower Upper
74 100
45 104.6 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.057 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 58099
Ion Ratio Lower Upper
101 100
85 47.1 37.0 55.6
151 71.4 58.6 87.8



#10
Methyl Iodide
Concen: 51.120 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion:142 Resp: 65198
Ion Ratio Lower Upper
142 100
127 47.6 41.6 62.4
141 13.8 11.6 17.4









#11
Tert butyl alcohol
Concen: 188.130 ug/l
RT: 2.953 min Scan# 307
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 59 Resp: 79525
Ion Ratio Lower Upper
59 100
57 10.2 6.0 9.0#



#12
1,1-Dichloroethene
Concen: 47.038 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 96 Resp: 54881
Ion Ratio Lower Upper
96 100
61 166.6 144.6 216.8
98 59.0 50.4 75.6





#13

Acrolein

Concen: 215.297 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 56 Resp: 56528

Ion Ratio Lower Upper

56 100

55 71.1 57.1 85.7



#14

Allyl chloride

Concen: 45.492 ug/l

RT: 2.660 min Scan# 258

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04




Tgt Ion: 41 Resp: 104466


Ion Ratio Lower Upper

41 100

39 69.3 55.8 83.6

76 31.5 24.3 36.5









#15
Acrylonitrile
Concen: 212.610 ug/l
RT: 3.056 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 185646
Ion Ratio Lower Upper
53 100
52 83.4 66.0 99.0
51 36.9 29.8 44.6



#16
Acetone
Concen: 235.539 ug/l
RT: 2.374 min Scan# 211
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 43 Resp: 215109
Ion Ratio Lower Upper
43 100
58 30.7 21.9 32.9









#17
Carbon Disulfide
Concen: 46.415 ug/l
RT: 2.508 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 155651
Ion Ratio Lower Upper
76 100
78 8.3 7.0 10.6



#18
Methyl Acetate
Concen: 45.321 ug/l
RT: 2.697 min Scan# 264
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 43 Resp: 86593
Ion Ratio Lower Upper
43 100
74 22.4 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 46.570 ug/l
RT: 3.105 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 192900
Ion Ratio Lower Upper
73 100
57 22.2 17.3 25.9



#20
Methylene Chloride
Concen: 45.293 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 84 Resp: 62568
Ion Ratio Lower Upper
84 100
49 126.2 103.6 155.4
51 38.7 33.8 50.6
86 61.7 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 48.139 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 59106
Ion Ratio Lower Upper
96 100
61 150.7 119.2 178.8
98 64.5 48.7 73.1



#22
Diisopropyl ether
Concen: 46.350 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 45 Resp: 203118
Ion Ratio Lower Upper
45 100
43 92.7 69.7 104.5
87 27.7 20.8 31.2
59 11.5 9.0 13.4









#23
Vinyl Acetate
Concen: 234.925 ug/l
RT: 3.715 min Scan# 413
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 956985
Ion Ratio Lower Upper
43 100
86 9.9 7.8 11.8



#24
1,1-Dichloroethane
Concen: 46.868 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 63 Resp: 112992
Ion Ratio Lower Upper
63 100
98 5.9 3.3 9.8
100 4.1 2.1 6.5









#25
2-Butanone
Concen: 214.185 ug/l
RT: 4.544 min Scan# 567
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 281101
Ion Ratio Lower Upper
43 100
72 24.0 19.4 29.2



#26
2,2-Dichloropropane
Concen: 48.556 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 77 Resp: 93953
Ion Ratio Lower Upper
77 100
97 22.3 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 48.646 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 96 Resp: 68505

Ion Ratio Lower Upper

96 100

61 150.9 0.0 312.2

98 64.3 0.0 129.0



#28

Bromochloromethane

Concen: 48.119 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04




Tgt Ion: 49 Resp: 52586


Ion Ratio Lower Upper

49 100

129 1.6 0.0 3.8

130 70.3 57.5 86.3





#29

Tetrahydrofuran

Concen: 206.438 ug/l

RT: 4.989 min Scan# 64

Delta R.T. -0.012 min

Lab File: VX039825.D


Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050






Tgt Ion: 42 Resp: 174385

Ion Ratio Lower Upper

42 100

72 43.8 34.2 51.4

71 40.4 32.2 48.2



#30

Chloroform

Concen: 49.247 ug/l

RT: 5.087 min Scan# 656

Delta R.T. -0.006 min

Lab File: VX039825.D




Acq: 19 Jan 2024 09:04


Tgt Ion: 83 Resp: 117206

Ion Ratio Lower Upper

83 100

85 64.5 50.3 75.5









#31
Cyclohexane
Concen: 46.477 ug/l
RT: 5.464 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	56	Resp:	99606
Ion	Ratio	Lower	Upper
56	100		
69	28.8	23.4	35.2
84	82.6	65.9	98.9



#32
1,1,1-Trichloroethane
Concen: 49.526 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion:	97	Resp:	100910
Ion	Ratio	Lower	Upper
97	100		
99	64.8	51.1	76.7
61	46.7	36.4	54.6





#33

1,2-Dichloroethane-d4

Concen: 46.670 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 65 Resp: 75787

Ion Ratio Lower Upper

65 100

67 50.9 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04




Tgt Ion: 114 Resp: 170299


Ion Ratio Lower Upper

114 100

63 22.1 0.0 45.0

88 16.3 0.0 36.6









#35
Dibromofluoromethane
Concen: 49.067 ug/l
RT: 5.379 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 56282
Ion Ratio Lower Upper
113 100
111 102.4 82.9 124.3
192 16.4 13.3 19.9



#36
1,1-Dichloropropene
Concen: 48.991 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 75 Resp: 87469
Ion Ratio Lower Upper
75 100
110 34.4 17.0 50.9
77 31.4 24.7 37.1





#37

Ethyl Acetate

Concen: 42.575 ug/l

RT: 4.702 min Scan# 594

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 43 Resp: 101408

Ion Ratio Lower Upper

43 100

61 13.9 10.6 16.0

70 10.2 8.3 12.5



#38

Carbon Tetrachloride

Concen: 51.880 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04




Tgt Ion: 117 Resp: 84836


Ion Ratio Lower Upper

117 100

119 94.0 78.2 117.2

121 31.5 24.4 36.6





#39

Methylcyclohexane

Concen: 48.842 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 104714

Ion Ratio Lower Upper

83 100

55 78.1 64.6 97.0

98 44.6 36.7 55.1



#40

Benzene

Concen: 48.140 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.006 min

Lab File: VX039825.D




Acq: 19 Jan 2024 09:04


Tgt Ion: 78 Resp: 243380

Ion Ratio Lower Upper

78 100

77 25.0 19.3 28.9









#41
Methacrylonitrile
Concen: 46.418 ug/l
RT: 4.910 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	41	Resp:	54066
Ion	Ratio	Lower	Upper
41	100		
39	56.4	45.2	67.8
67	72.2	52.6	78.8
52	31.9	22.7	34.1



#42
1,2-Dichloroethane
Concen: 50.278 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion:	62	Resp:	96637
Ion	Ratio	Lower	Upper
62	100		
98	8.1	0.0	16.0









#43
Isopropyl Acetate
Concen: 47.358 ug/l
RT: 6.330 min Scan# 860
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 165472
Ion Ratio Lower Upper
43 100
61 18.5 15.2 22.8
87 12.6 9.8 14.6



#44
Trichloroethene
Concen: 50.768 ug/l
RT: 7.117 min Scan# 989
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion:130 Resp: 62903
Ion Ratio Lower Upper
130 100
95 103.7 0.0 204.4





#45

1,2-Dichloropropane

Concen: 47.773 ug/l

RT: 7.421 min Scan# 1039

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 63 Resp: 64791

Ion Ratio Lower Upper

63 100

65 30.3 24.3 36.5



#46

Dibromomethane

Concen: 50.818 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04




Tgt Ion: 93 Resp: 47603


Ion Ratio Lower Upper

93 100

95 81.0 66.6 100.0

174 83.7 69.4 104.0









#47
Bromodichloromethane
Concen: 52.325 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 90779
Ion Ratio Lower Upper
83 100
85 67.3 49.8 74.6
127 8.3 7.1 10.7



#48
Methyl methacrylate
Concen: 46.995 ug/l
RT: 7.684 min Scan# 1082
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 41 Resp: 80712
Ion Ratio Lower Upper
41 100
69 79.6 59.2 88.8
39 64.9 47.7 71.5









#49
1,4-Dioxane
Concen: 813.181 ug/l
RT: 7.653 min Scan# 1077
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 88 Resp: 27024
Ion Ratio Lower Upper
88 100
43 37.0 0.0 0.0#
58 68.6 60.1 90.1



#50
Toluene-d8
Concen: 48.627 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 98 Resp: 210366
Ion Ratio Lower Upper
98 100
100 64.0 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 227.521 ug/l
RT: 8.568 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 547512
Ion Ratio Lower Upper
43 100
58 36.9 30.4 45.6



#52
Toluene
Concen: 50.757 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 92 Resp: 153270
Ion Ratio Lower Upper
92 100
91 176.1 134.2 201.2









#53
t-1,3-Dichloropropene
Concen: 54.003 ug/l
RT: 8.976 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 99415
Ion Ratio Lower Upper
75 100
77 31.5 25.7 38.5



#54
cis-1,3-Dichloropropene
Concen: 52.133 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 75 Resp: 105529
Ion Ratio Lower Upper
75 100
77 32.1 25.3 37.9
39 56.2 46.7 70.1









#55
1,1,2-Trichloroethane
Concen: 49.627 ug/l
RT: 9.147 min Scan# 1316
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 97	Resp: 63803
Ion Ratio	Lower Upper
97 100	
83 84.9	65.8 98.8
85 58.0	41.6 62.4
99 60.6	47.8 71.8



#56
Ethyl methacrylate
Concen: 49.867 ug/l
RT: 9.110 min Scan# 1316
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 69	Resp: 103764
Ion Ratio	Lower Upper
69 100	
41 72.2	56.8 85.2
39 44.7	32.0 48.0





#57

1,3-Dichloropropane

Concen: 49.135 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 76 Resp: 107053

Ion Ratio Lower Upper

76 100

78 32.4 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 235.359 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX039825.D




Acq: 19 Jan 2024 09:04


Tgt Ion: 63 Resp: 231486

Ion Ratio Lower Upper

63 100

106 26.0 22.5 33.7





#59

2-Hexanone

Concen: 234.026 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 43 Resp: 445716

Ion Ratio Lower Upper

43 100

58 51.1 26.2 78.5



#60

Dibromochloromethane

Concen: 55.829 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX039825.D




Acq: 19 Jan 2024 09:04


Tgt Ion: 129 Resp: 67690

Ion Ratio Lower Upper

129 100

127 77.8 38.8 116.4









#61
1,2-Dibromoethane
Concen: 51.659 ug/l
RT: 9.604 min Scan# 1398
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 68777
Ion Ratio Lower Upper
107 100
109 93.9 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 50.677 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 95 Resp: 85731
Ion Ratio Lower Upper
95 100
174 67.2 0.0 135.6
176 64.7 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 157761
Ion Ratio Lower Upper
117 100
82 61.2 46.2 69.4
119 32.4 25.5 38.3



#64
Tetrachloroethene
Concen: 51.144 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion:164 Resp: 52470
Ion Ratio Lower Upper
164 100
166 129.7 102.7 154.1
129 105.9 80.4 120.6
131 100.7 76.5 114.7









#65
Chlorobenzene
Concen: 49.621 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 163045
Ion Ratio Lower Upper
112 100
114 32.2 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 53.165 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion:131 Resp: 59483
Ion Ratio Lower Upper
131 100
133 95.7 47.1 141.4
119 66.2 32.2 96.6









#67
Ethyl Benzene
Concen: 50.594 ug/l
RT: 10.189 min Scan# 1494
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 308638
Ion Ratio Lower Upper
91 100
106 30.0 24.9 37.3



#68
m/p-Xylenes
Concen: 100.782 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion:106 Resp: 226104
Ion Ratio Lower Upper
106 100
91 215.9 165.5 248.3





#69

o-Xylene

Concen: 49.964 ug/l

RT: 10.640 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion:106 Resp: 111008

Ion Ratio Lower Upper

106 100

91 226.8 108.9 326.8



#70

Styrene

Concen: 51.313 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04




Tgt Ion:104 Resp: 186233


Ion Ratio Lower Upper

104 100

78 56.5 43.6 65.4

103 55.6 43.6 65.4





#71

Bromoform

Concen: 55.890 ug/l

RT: 10.799 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:173 Resp: 46954

Ion Ratio Lower Upper

173 100

175 47.5 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04




Tgt Ion:152 Resp: 77864


Ion Ratio Lower Upper

152 100

115 96.0 43.3 129.9

150 176.0 0.0 345.0





#73

Isopropylbenzene

Concen: 48.803 ug/l

RT: 10.957 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion:105 Resp: 302540

Ion Ratio Lower Upper

105 100

120 25.8 12.9 38.6



#74

N-ethyl acetate

Concen: 47.638 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Tgt Ion: 43 Resp: 153938




Ion Ratio Lower Upper


43 100

70 38.2 32.9 49.3

55 22.8 19.3 28.9

61 22.3 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 45.737 ug/l

RT: 11.207 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 108269

Ion Ratio Lower Upper

83 100

131 9.1 5.3 15.8

85 63.2 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 59.434 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX039825.D




Acq: 19 Jan 2024 09:04


Tgt Ion: 75 Resp: 119980

Ion Ratio Lower Upper

75 100

77 32.6 19.1 57.5





#77

Bromobenzene

Concen: 49.266 ug/l

RT: 11.195 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:156 Resp: 68001

Ion Ratio Lower Upper

156 100

77 180.9 88.4 265.3

158 95.3 48.0 144.0



#78

n-propylbenzene

Concen: 49.975 ug/l

RT: 11.299 min Scan# 1675

Delta R.T. -0.006 min

Lab File: VX039825.D




Acq: 19 Jan 2024 09:04


Tgt Ion: 91 Resp: 374974

Ion Ratio Lower Upper

91 100

120 21.5 11.5 34.4









#79
2-Chlorotoluene
Concen: 47.423 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 219920
Ion Ratio Lower Upper
91 100
126 31.3 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 49.589 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 105 Resp: 261094
Ion Ratio Lower Upper
105 100
120 45.4 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 47.556 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 31978
Ion Ratio Lower Upper
75 100
53 107.8 90.5 135.7
89 45.4 36.9 55.3



#82
4-Chlorotoluene
Concen: 49.382 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 91 Resp: 259757
Ion Ratio Lower Upper
91 100
126 27.4 14.2 42.6





#83

tert-Butylbenzene

Concen: 49.869 ug/l

RT: 11.713 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:119 Resp: 250675

Ion Ratio Lower Upper

119 100

91 64.2 31.2 93.6

134 21.9 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 50.230 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX039825.D




Acq: 19 Jan 2024 09:04


Tgt Ion:105 Resp: 263851

Ion Ratio Lower Upper

105 100

120 42.4 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 51.381 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX039825.D
 Acq: 19 Jan 2024 09:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 335269
 Ion Ratio Lower Upper
 105 100
 134 19.0 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 51.902 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX039825.D
 Acq: 19 Jan 2024 09:04

Tgt Ion:119 Resp: 270673
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.8 38.3
 91 28.4 13.1 39.3









#87
1,3-Dichlorobenzene
Concen: 50.287 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 133053
Ion Ratio Lower Upper
146 100
111 45.2 22.1 66.1
148 64.2 31.8 95.3



#88
1,4-Dichlorobenzene
Concen: 48.748 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.006 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion:146 Resp: 133214
Ion Ratio Lower Upper
146 100
111 44.2 21.3 63.7
148 63.8 31.8 95.4









#89
n-Butylbenzene
Concen: 54.290 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 273457
Ion Ratio Lower Upper
91 100
92 51.1 27.4 82.0
134 23.2 12.6 37.8



#90
Hexachloroethane
Concen: 53.587 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX039825.D
Acq: 19 Jan 2024 09:04

Tgt Ion: 117 Resp: 47146
Ion Ratio Lower Upper
117 100
201 64.8 32.9 98.6









#91
 1,2-Dichlorobenzene
 Concen: 48.855 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX039825.D
 Acq: 19 Jan 2024 09:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:146 Resp: 127792
 Ion Ratio Lower Upper
 146 100
 111 46.2 22.4 67.0
 148 63.3 31.6 95.0



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 46.889 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. -0.006 min
 Lab File: VX039825.D
 Acq: 19 Jan 2024 09:04

Tgt Ion: 75 Resp: 27199
 Ion Ratio Lower Upper
 75 100
 155 69.3 36.1 108.5
 157 88.7 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 52.248 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:180 Resp: 83520

Ion Ratio Lower Upper

180 100

182 94.8 48.3 144.8

145 34.4 18.0 54.0



#94

Hexachlorobutadiene

Concen: 55.537 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04




Tgt Ion:225 Resp: 36904


Ion Ratio Lower Upper

225 100

223 63.5 31.1 93.2

227 62.2 32.5 97.5





#95

Naphthalene

Concen: 48.115 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:128 Resp: 283144

Ion Ratio Lower Upper

128 100

127 12.9 10.3 15.5

129 10.8 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 49.671 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.006 min

Lab File: VX039825.D

Acq: 19 Jan 2024 09:04

Tgt Ion:180 Resp: 80135

Ion Ratio Lower Upper

180 100

182 95.7 46.9 140.6

145 37.0 17.8 53.5

