

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039827.D
 Acq On : 19 Jan 2024 10:08
 Operator : JC/MD
 Sample : VX0119WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0119WBL01

Quant Time: Jan 20 00:20:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	99127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	180144	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83249	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76781	46.239	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.480%
35) Dibromofluoromethane	5.385	113	59745	49.240	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.480%
50) Toluene-d8	8.647	98	221528	48.409	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.820%
62) 4-Bromofluorobenzene	11.079	95	85417	47.732	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039827.D
Acq On : 19 Jan 2024 10:08
Operator : JC/MD
Sample : VX0119WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0119WBL01

Quant Time: Jan 20 00:20:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

