

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039840.D
 Acq On : 19 Jan 2024 15:36
 Operator : JC/MD
 Sample : P1184-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUMOGW-4-10-13

Quant Time: Jan 20 00:27:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	83791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	154963	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	149911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	72938	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65882	46.937	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.880%
35) Dibromofluoromethane	5.385	113	50852	48.721	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	8.647	98	191357	48.610	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.220%
62) 4-Bromofluorobenzene	11.079	95	75046	48.751	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.500%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1982	2.511	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	1209	0.993	ug/l	76
44) Trichloroethene	7.123	130	1758	1.559	ug/l	85
64) Tetrachloroethene	9.275	164	1383	1.419	ug/l	89
93) 1,2,4-Trichlorobenzene	13.585	180	5294	3.535	ug/l	94
96) 1,2,3-Trichlorobenzene	13.963	180	1131	0.748	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039840.D
Acq On : 19 Jan 2024 15:36
Operator : JC/MD
Sample : P1184-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-4-10-13

Quant Time: Jan 20 00:27:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

