

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039843.D
 Acq On : 19 Jan 2024 16:44
 Operator : JC/MD
 Sample : P1184-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUMOGW-5-10-13

Quant Time: Jan 20 00:28:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	92199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	171325	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163502	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	82171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73918	47.860	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.720%
35) Dibromofluoromethane	5.385	113	56538	48.995	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.000%
50) Toluene-d8	8.647	98	210713	48.415	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.840%
62) 4-Bromofluorobenzene	11.079	95	84858	49.860	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.720%
Target Compounds						
					Qvalue	
30) Chloroform	5.093	83	2020	0.892	ug/l	98
44) Trichloroethene	7.129	130	2149	1.724	ug/l	81
64) Tetrachloroethene	9.269	164	1655	1.557	ug/l #	82
93) 1,2,4-Trichlorobenzene	13.585	180	2278	1.350	ug/l	96
96) 1,2,3-Trichlorobenzene	13.957	180	473	0.278	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039843.D
Acq On : 19 Jan 2024 16:44
Operator : JC/MD
Sample : P1184-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-5-10-13

Quant Time: Jan 20 00:28:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

