

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039847.D
 Acq On : 19 Jan 2024 18:15
 Operator : JC/MD
 Sample : P1196-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20240118

Quant Time: Jan 20 03:42:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	148595	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	145101	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63052	46.483	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.960%
35) Dibromofluoromethane	5.385	113	48023	47.982	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.960%
50) Toluene-d8	8.647	98	182791	48.424	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.840%
62) 4-Bromofluorobenzene	11.079	95	71289	48.295	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.580%
Target Compounds						
64) Tetrachloroethene	9.269	164	941	0.997	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039847.D
Acq On : 19 Jan 2024 18:15
Operator : JC/MD
Sample : P1196-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20240118

Quant Time: Jan 20 03:42:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

