

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039853.D
 Acq On : 19 Jan 2024 21:16
 Operator : JC/MD
 Sample : VX0119MBL02
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0119MBL02

Quant Time: Jan 20 03:25:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	148379	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141144	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66890	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62944	45.920	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.840%
35) Dibromofluoromethane	5.379	113	49334	49.364	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.720%
50) Toluene-d8	8.647	98	181268	48.091	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.180%
62) 4-Bromofluorobenzene	11.079	95	70054	47.527	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039853.D
Acq On : 19 Jan 2024 21:16
Operator : JC/MD
Sample : VX0119MBL02
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0119MBL02

Quant Time: Jan 20 03:25:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

