

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039858.D
 Acq On : 19 Jan 2024 23:09
 Operator : JC/MD
 Sample : PB158526ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158526ZHE#01

Quant Time: Jan 20 03:54:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	149633	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	147442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70769	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63949	46.289	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.580%
35) Dibromofluoromethane	5.385	113	48833	48.453	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.900%
50) Toluene-d8	8.647	98	188030	49.467	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.940%
62) 4-Bromofluorobenzene	11.079	95	73134	49.201	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.400%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	18434	23.726	ug/l	90
20) Methylene Chloride	2.794	84	3811	3.243	ug/l	96
37) Ethyl Acetate	4.739	43	2452	1.172	ug/l #	75
43) Isopropyl Acetate	6.336	43	80080	26.084	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039858.D
Acq On : 19 Jan 2024 23:09
Operator : JC/MD
Sample : PB158526ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158526ZHE#01

Quant Time: Jan 20 03:54:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

