

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039859.D
 Acq On : 19 Jan 2024 23:31
 Operator : JC/MD
 Sample : PB158526ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158526ZHE#02

Quant Time: Jan 20 03:54:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	94407	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	170722	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167565	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81063	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72655	45.942	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.880%
35) Dibromofluoromethane	5.385	113	55906	48.619	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.240%
50) Toluene-d8	8.646	98	215129	49.605	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.200%
62) 4-Bromofluorobenzene	11.079	95	83819	49.424	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.840%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	19133	21.512	ug/l	95
20) Methylene Chloride	2.788	84	3799	2.824	ug/l #	92
43) Isopropyl Acetate	6.336	43	78961	22.542	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039859.D
Acq On : 19 Jan 2024 23:31
Operator : JC/MD
Sample : PB158526ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158526ZHE#02

Quant Time: Jan 20 03:54:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

