

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039863.D
 Acq On : 20 Jan 2024 01:02
 Operator : JC/MD
 Sample : PB158526ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158526ZHE#06

Quant Time: Jan 20 03:56:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	84914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	156621	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	74938	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68268	47.994	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.980%
35) Dibromofluoromethane	5.385	113	52224	49.506	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.020%
50) Toluene-d8	8.647	98	197025	49.520	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.040%
62) 4-Bromofluorobenzene	11.079	95	79042	50.803	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.600%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	19367	24.210	ug/l	95
20) Methylene Chloride	2.788	84	3888	3.213	ug/l	92
43) Isopropyl Acetate	6.336	43	79603	24.772	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039863.D
Acq On : 20 Jan 2024 01:02
Operator : JC/MD
Sample : PB158526ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158526ZHE#06

Quant Time: Jan 20 03:56:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

