

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
 Data File : VX039865.D
 Acq On : 20 Jan 2024 01:48
 Operator : JC/MD
 Sample : PB158526ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158526ZHE#08

Quant Time: Jan 20 03:57:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	151154	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147934	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63610	46.849	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.700%
35) Dibromofluoromethane	5.385	113	48906	48.037	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.080%
50) Toluene-d8	8.647	98	190128	49.515	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.040%
62) 4-Bromofluorobenzene	11.079	95	73461	48.924	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.840%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	13209	17.298	ug/l	96
20) Methylene Chloride	2.788	84	3619	3.133	ug/l	94
25) 2-Butanone	4.556	43	6048	5.512	ug/l	91
43) Isopropyl Acetate	6.336	43	72748	23.457	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039865.D
Acq On : 20 Jan 2024 01:48
Operator : JC/MD
Sample : PB158526ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158526ZHE#08

Quant Time: Jan 20 03:57:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

