



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039824.D
Acq On : 19 Jan 2024 08:00
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Title : SW846 8260
Last Update : Thu Jan 11 03:51:26 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.5	6894	PASS
75	95	30	60	54.3	18286	PASS
95	95	100	100	100.0	33669	PASS
96	95	5	9	6.7	2241	PASS
173	174	0.00	2	1.0	222	PASS
174	95	50	100	68.9	23187	PASS
175	174	5	9	6.7	1547	PASS
176	174	95	101	96.6	22395	PASS
177	176	5	9	6.4	1434	PASS