



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011924\
Data File : VX039851.D
Acq On : 19 Jan 2024 20:08
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Title : SW846 8260
Last Update : Thu Jan 11 03:51:26 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.1	5896	PASS
75	95	30	60	55.3	15454	PASS
95	95	100	100	100.0	27928	PASS
96	95	5	9	6.7	1863	PASS
173	174	0.00	2	0.7	131	PASS
174	95	50	100	67.9	18950	PASS
175	174	5	9	7.7	1451	PASS
176	174	95	101	95.0	17994	FAIL*
177	176	5	9	6.6	1180	PASS