

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012020\
 Data File : VX014668.D
 Acq On : 20 Jan 2020 12:01
 Operator : JC/SP
 Sample : L1202-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 F1-F2-(SETTLED)

Quant Time: Jan 21 00:48:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	470747	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	723584	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	654434	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	285736	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	264609	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
35) Dibromofluoromethane	5.48	113	224854	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
50) Toluene-d8	8.71	98	864167	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
62) 4-Bromofluorobenzene	11.13	95	283094	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
Target Compounds						
30) Chloroform	5.20	83	151959	18.122	ug/l	97
47) Bromodichloromethane	7.89	83	5601	0.857	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012020\
Data File : VX014668.D
Acq On : 20 Jan 2020 12:01
Operator : JC/SP
Sample : L1202-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
F1-F2-(SETTLED)

Quant Time: Jan 21 00:48:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

