


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX012020\
 Data File : VX014663.D
 Acq On : 20 Jan 2020 08:44
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Title : SW846 8260
 Last Update : Tue Jan 07 15:54:41 2020



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1670

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	9399	PASS
75	95	30	60	51.1	25571	PASS
95	95	100	100	100.0	50051	PASS
96	95	5	9	6.6	3290	PASS
173	174	0.00	2	1.1	457	PASS
174	95	50	100	83.3	41682	PASS
175	174	5	9	7.4	3100	PASS
176	174	95	101	96.0	40002	PASS
177	176	5	9	6.6	2628	PASS