

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX012021\  
 Data File : VX020676.D  
 Acq On : 20 Jan 2021 11:45  
 Operator : JC/SP  
 Sample : VX0120MBL01  
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0120MBL01

Quant Time: Jan 20 16:55:19 2021  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X011921W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 19 14:09:08 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.623  | 168   | 249990   | 50.00    | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.824  | 114   | 428533   | 50.00    | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.092 | 117   | 399053   | 50.00    | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.061 | 152   | 180463   | 50.00    | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 6.025  | 65    | 165967   | 48.56    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =     | 97.12%   |
| 35) Dibromofluoromethane    | 5.458  | 113   | 140449   | 47.57    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =     | 95.14%   |
| 50) Toluene-d8              | 8.690  | 98    | 517822   | 48.64    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =     | 97.28%   |
| 62) 4-Bromofluorobenzene    | 11.122 | 95    | 191634   | 47.62    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =     | 95.24%   |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX012021\  
Data File : VX020676.D  
Acq On : 20 Jan 2021 11:45  
Operator : JC/SP  
Sample : VX0120MBL01  
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0120MBL01

Quant Time: Jan 20 16:55:19 2021  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X011921W.M  
Quant Title : SW846 8260  
QLast Update : Tue Jan 19 14:09:08 2021  
Response via : Initial Calibration

