

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012021\
 Data File : VX020678.D
 Acq On : 20 Jan 2021 13:24
 Operator : JC/SP
 Sample : M1178-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-011921

Quant Time: Jan 20 16:56:03 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 14:09:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	245475	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	434125	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	405325	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	179271	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	167976	50.05	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.10%
35) Dibromofluoromethane	5.458	113	141830	47.42	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.84%
50) Toluene-d8	8.689	98	529975	49.14	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.28%
62) 4-Bromofluorobenzene	11.122	95	194664	47.75	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.50%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012021\
Data File : VX020678.D
Acq On : 20 Jan 2021 13:24
Operator : JC/SP
Sample : M1178-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-011921

Quant Time: Jan 20 16:56:03 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 14:09:08 2021
Response via : Initial Calibration

