

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012021\
 Data File : VX020680.D
 Acq On : 20 Jan 2021 14:11
 Operator : JC/SP
 Sample : M1161-01
 Misc : 6.50g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-05-011521

Manual Integrations
 APPROVED

MMDadoda
 1/21/2021 6:36:38 PM

Quant Time: Jan 20 16:56:50 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 14:09:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	257223	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	429463	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	385337	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	175353	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	163597	45.46	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	90.92%		
35) Dibromofluoromethane	5.464	113	136377	45.00	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	90.00%		
50) Toluene-d8	8.696	98	513338	48.39	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.78%		
62) 4-Bromofluorobenzene	11.122	95	183418	47.07	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.14%		
Target Compounds					Qvalue	
18) Methyl Acetate	2.764	43	29879	9.63	ug/l	98
20) Methylene Chloride	2.825	84	5907	1.78	ug/l	97
40) Benzene	6.105	78	10494	0.83	ug/l	95
44) Trichloroethene	7.184	130	43254	12.53	ug/l	95
52) Toluene	8.763	92	6279	0.81	ug/l	100
67) Ethyl Benzene	10.232	91	6660	0.47	ug/l	96
68) m/p-Xylenes	10.342	106	6832	1.30	ug/l	96
69) o-Xylene	10.683	106	2920	0.58	ug/l	99
78) n-propylbenzene	11.341	91	6064	0.43	ug/l	99
80) 1,3,5-Trimethylbenzene	11.488	105	7802	0.75	ug/l	99
84) 1,2,4-Trimethylbenzene	11.793	105	29565	2.98	ug/l	98
85) sec-Butylbenzene	11.933	105	3627	0.31	ug/l	84
86) p-Isopropyltoluene	12.049	119	2739m	1.11	ug/l	
89) n-Butylbenzene	12.366	91	9639	2.32	ug/l #	51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012021\
Data File : VX020680.D
Acq On : 20 Jan 2021 14:11
Operator : JC/SP
Sample : M1161-01
Misc : 6.50g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-05-011521

Manual Integrations
APPROVED

MMDadoda
1/21/2021 6:36:38 PM

Quant Time: Jan 20 16:56:50 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 14:09:08 2021
Response via : Initial Calibration

