

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012021\
 Data File : VX020681.D
 Acq On : 20 Jan 2021 14:59
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 20 16:57:15 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 14:09:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	263034	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	446313	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	410464	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	180494	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	171795	46.69	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.38%
35) Dibromofluoromethane	5.458	113	146934	46.66	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.32%
50) Toluene-d8	8.689	98	550532	49.93	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.86%
62) 4-Bromofluorobenzene	11.116	95	195926	48.38	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.76%
Target Compounds						
44) Trichloroethene	7.184	130	5709	1.59	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012021\
Data File : VX020681.D
Acq On : 20 Jan 2021 14:59
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 20 16:57:15 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 14:09:08 2021
Response via : Initial Calibration

