

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012022\
 Data File : VX026523.D
 Acq On : 20 Jan 2022 11:04
 Operator : JC/MD
 Sample : VX0120MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0120MBL01

Quant Time: Jan 21 04:01:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	138783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	231610	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81859	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92504	47.715	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.420%
35) Dibromofluoromethane	5.391	113	73245	48.972	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.940%
50) Toluene-d8	8.647	98	265588	49.277	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.560%
62) 4-Bromofluorobenzene	11.079	95	91810	47.952	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012022\
Data File : VX026523.D
Acq On : 20 Jan 2022 11:04
Operator : JC/MD
Sample : VX0120MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0120MBL01

Quant Time: Jan 21 04:01:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

