

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012022\
 Data File : VX026528.D
 Acq On : 20 Jan 2022 13:11
 Operator : JC/MD
 Sample : PB142135ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142135ZHE#02

Quant Time: Jan 21 04:03:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	133740	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	227631	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83008	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91369	48.906	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.820%
35) Dibromofluoromethane	5.391	113	71555	48.678	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.360%
50) Toluene-d8	8.653	98	267820	50.559	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.120%
62) 4-Bromofluorobenzene	11.085	95	94138	50.027	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.060%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5991	6.288	ug/l	99
20) Methylene Chloride	2.794	84	3315	2.106	ug/l #	93
43) Isopropyl Acetate	6.348	43	11552	3.046	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012022\
Data File : VX026528.D
Acq On : 20 Jan 2022 13:11
Operator : JC/MD
Sample : PB142135ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142135ZHE#02

Quant Time: Jan 21 04:03:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

