

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012022\
 Data File : VX026529.D
 Acq On : 20 Jan 2022 13:34
 Operator : JC/MD
 Sample : PB142135ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142135ZHE#03

Quant Time: Jan 21 04:04:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	133556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227454	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84321	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	91919	49.268	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.540%		
35) Dibromofluoromethane	5.391	113	72332	49.245	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.480%		
50) Toluene-d8	8.652	98	265927	50.241	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	100.480%		
62) 4-Bromofluorobenzene	11.085	95	94295	50.149	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.300%		
Target Compounds						
16) Acetone	2.385	43	5790	6.086	ug/l	98
20) Methylene Chloride	2.794	84	3416	2.173	ug/l	97
43) Isopropyl Acetate	6.348	43	11882	3.136	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012022\
Data File : VX026529.D
Acq On : 20 Jan 2022 13:34
Operator : JC/MD
Sample : PB142135ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142135ZHE#03

Quant Time: Jan 21 04:04:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

