

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012022\
 Data File : VX026530.D
 Acq On : 20 Jan 2022 13:58
 Operator : JC/MD
 Sample : PB142135ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142135ZHE#04

Quant Time: Jan 21 04:04:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133485	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229289	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82063	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92053	49.367	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.740%
35) Dibromofluoromethane	5.391	113	71922	48.574	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.140%
50) Toluene-d8	8.653	98	265032	49.671	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.340%
62) 4-Bromofluorobenzene	11.079	95	92796	48.957	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.920%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	5872	6.175	ug/l	97
20) Methylene Chloride	2.794	84	3268	2.080	ug/l	98
43) Isopropyl Acetate	6.348	43	12236	3.204	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012022\
Data File : VX026530.D
Acq On : 20 Jan 2022 13:58
Operator : JC/MD
Sample : PB142135ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142135ZHE#04

Quant Time: Jan 21 04:04:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

