

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012023\
 Data File : VX033822.D
 Acq On : 20 Jan 2023 12:48
 Operator : JC/MD
 Sample : 01242-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10-GW-148-150

Quant Time: Jan 21 04:43:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	171500	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	249648	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105142	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83455	46.217	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.440%
35) Dibromofluoromethane	5.385	113	74812	48.260	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.520%
50) Toluene-d8	8.647	98	283531	50.199	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.400%
62) 4-Bromofluorobenzene	11.079	95	97470	46.480	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.960%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	3.117	73	4947	1.021	ug/l	100
22) Diisopropyl ether	3.757	45	1379	0.260	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012023\
Data File : VX033822.D
Acq On : 20 Jan 2023 12:48
Operator : JC/MD
Sample : 01242-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-GW-148-150

Quant Time: Jan 21 04:43:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

