

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012023\
 Data File : VX033832.D
 Acq On : 20 Jan 2023 16:40
 Operator : JC/MD
 Sample : 01242-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10-GW-58-60

Quant Time: Jan 21 04:46:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	165087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	243599	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215199	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101036	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82094	47.229	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.460%
35) Dibromofluoromethane	5.385	113	74324	49.136	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.280%
50) Toluene-d8	8.647	98	275409	49.972	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.940%
62) 4-Bromofluorobenzene	11.079	95	94954	46.404	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.800%
Target Compounds						
16) Acetone	2.386	43	8771	1.574	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012023\
Data File : VX033832.D
Acq On : 20 Jan 2023 16:40
Operator : JC/MD
Sample : 01242-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-GW-58-60

Quant Time: Jan 21 04:46:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

