

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012119\
 Data File : VX007162.D
 Acq On : 21 Jan 2019 16:41
 Operator : JC/SP
 Sample : K1198-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW518-3337

Quant Time: Jan 22 10:54:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	541690	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	833396	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	791116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	393168	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	290655	51.34	ug/l	0.00
Spiked Amount						
			Recovery	=	102.68%	
35) Dibromofluoromethane	5.50	113	257141	48.18	ug/l	0.00
Spiked Amount						
			Recovery	=	96.36%	
50) Toluene-d8	8.71	98	999119	51.14	ug/l	0.00
Spiked Amount						
			Recovery	=	102.28%	
62) 4-Bromofluorobenzene	11.13	95	360981	53.00	ug/l	0.00
Spiked Amount						
			Recovery	=	106.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012119\
Data File : VX007162.D
Acq On : 21 Jan 2019 16:41
Operator : JC/SP
Sample : K1198-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW518-3337

Quant Time: Jan 22 10:54:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

