

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012119\
 Data File : VX007163.D
 Acq On : 21 Jan 2019 17:08
 Operator : JC/SP
 Sample : K1199-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW528-2327

Quant Time: Jan 22 01:22:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	532393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	831378	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	781997	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	395273	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	289513	52.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.06%	
35) Dibromofluoromethane	5.50	113	256545	48.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.38%	
50) Toluene-d8	8.71	98	993966	51.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.00%	
62) 4-Bromofluorobenzene	11.13	95	359520	52.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.84%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012119\
Data File : VX007163.D
Acq On : 21 Jan 2019 17:08
Operator : JC/SP
Sample : K1199-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW528-2327

Quant Time: Jan 22 01:22:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

