

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012119\
 Data File : VX007167.D
 Acq On : 21 Jan 2019 18:54
 Operator : JC/SP
 Sample : K1195-04DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW036-190116DL

Quant Time: Jan 22 10:56:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	511076	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	795490	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	742836	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	381302	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	275230	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
35) Dibromofluoromethane	5.50	113	242849	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
50) Toluene-d8	8.71	98	959656	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
62) 4-Bromofluorobenzene	11.13	95	342236	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
Target Compounds						
65) Chlorobenzene	10.13	112	638222	39.392	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012119\
Data File : VX007167.D
Acq On : 21 Jan 2019 18:54
Operator : JC/SP
Sample : K1195-04DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW036-190116DL

Quant Time: Jan 22 10:56:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

