

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012119\
 Data File : VX007168.D
 Acq On : 21 Jan 2019 19:20
 Operator : JC/SP
 Sample : K1195-05DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS038MW013-190116DL

Quant Time: Jan 22 02:07:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	566581	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	875103	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	833056	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	426774	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	304601	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
35) Dibromofluoromethane	5.51	113	271202	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
50) Toluene-d8	8.71	98	1055287	51.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.90%	
62) 4-Bromofluorobenzene	11.13	95	386214	54.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.02%	
Target Compounds						
40) Benzene	6.15	78	14798	0.637	ug/l	96
65) Chlorobenzene	10.13	112	549813	30.260	ug/l	98
73) Isopropylbenzene	11.02	105	32561	1.082	ug/l	100
78) n-propylbenzene	11.36	91	26189	0.793	ug/l	99
83) tert-Butylbenzene	11.77	119	28042	1.118	ug/l	87
85) sec-Butylbenzene	11.94	105	50980	1.701	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012119\
Data File : VX007168.D
Acq On : 21 Jan 2019 19:20
Operator : JC/SP
Sample : K1195-05DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS038MW013-190116DL

Quant Time: Jan 22 02:07:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

