

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012120\
 Data File : VX014674.D
 Acq On : 21 Jan 2020 10:12
 Operator : JC/SP
 Sample : VX0121WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0121WBL01

Quant Time: Jan 22 04:37:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	419299	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	656020	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	589641	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	258405	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	247439	52.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.72%	
35) Dibromofluoromethane	5.48	113	201281	50.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.38%	
50) Toluene-d8	8.71	98	769417	49.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.86%	
62) 4-Bromofluorobenzene	11.13	95	258036	45.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.22%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012120\
Data File : VX014674.D
Acq On : 21 Jan 2020 10:12
Operator : JC/SP
Sample : VX0121WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0121WBL01

Quant Time: Jan 22 04:37:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

