


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX012120\
Data File : VX014672.D
Acq On : 21 Jan 2020 09:03
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Title : SW846 8260
Last Update : Tue Jan 07 15:54:41 2020



AutoFind: Scans 1680, 1681, 1682; Background Corrected with Scan 1669

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	12865	PASS
75	95	30	60	50.5	33689	PASS
95	95	100	100	100.0	66733	PASS
96	95	5	9	6.6	4400	PASS
173	174	0.00	2	1.5	908	PASS
174	95	50	100	90.9	60658	PASS
175	174	5	9	8.0	4859	PASS
176	174	95	101	95.5	57936	PASS
177	176	5	9	6.5	3738	PASS