

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012122\
 Data File : VX026539.D
 Acq On : 21 Jan 2022 11:34
 Operator : JC/MD
 Sample : VX0121WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0121WBL01

Quant Time: Jan 22 03:57:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	135560	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91552	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91567	48.354	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.700%
35) Dibromofluoromethane	5.391	113	71657	48.500	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.000%
50) Toluene-d8	8.653	98	270381	50.784	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.560%
62) 4-Bromofluorobenzene	11.079	95	102552	54.222	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012122\
Data File : VX026539.D
Acq On : 21 Jan 2022 11:34
Operator : JC/MD
Sample : VX0121WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0121WBL01

Quant Time: Jan 22 03:57:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

