

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012122\
 Data File : VX026544.D
 Acq On : 21 Jan 2022 13:40
 Operator : JC/MD
 Sample : N1229-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 22 03:58:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	143243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	240566	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94438	47.195	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.400%
35) Dibromofluoromethane	5.391	113	75255	48.442	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.880%
50) Toluene-d8	8.653	98	284933	50.898	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.800%
62) 4-Bromofluorobenzene	11.079	95	101923	51.252	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012122\
Data File : VX026544.D
Acq On : 21 Jan 2022 13:40
Operator : JC/MD
Sample : N1229-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 22 03:58:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

