

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX012122\
 Data File : VX026537.D
 Acq On : 21 Jan 2022 10:37
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 22 03:56:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	50.000	45.921	8.2	87	0.00
3 P	Chloromethane	50.000	50.285	-0.6	98	0.00
4 C	Vinyl Chloride	50.000	48.924	2.2#	91	0.00
5 T	Bromomethane	50.000	54.102	-8.2	105	0.00
6 T	Chloroethane	50.000	46.421	7.2	89	0.00
7 T	Trichlorofluoromethane	50.000	46.937	6.1	86	0.00
8 T	Diethyl Ether	50.000	48.550	2.9	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.966	2.1	91	0.00
10 T	Methyl Iodide	50.000	58.923	-17.8	118	0.00
11 T	Tert butyl alcohol	250.000	237.077	5.2	86	0.00
12 CM	1,1-Dichloroethene	50.000	47.985	4.0#	89	0.00
13 T	Acrolein	250.000	252.335	-0.9	101	0.00
14 T	Allyl chloride	50.000	48.600	2.8	88	0.00
15 T	Acrylonitrile	250.000	259.677	-3.9	98	0.00
16 T	Acetone	250.000	248.187	0.7	92	0.00
17 T	Carbon Disulfide	50.000	44.572	10.9	85	0.00
18 T	Methyl Acetate	50.000	50.554	-1.1	97	0.00
19 T	Methyl tert-butyl Ether	50.000	47.250	5.5	88	0.00
20 T	Methylene Chloride	50.000	48.744	2.5	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.162	3.7	90	0.00
22 T	Diisopropyl ether	50.000	49.051	1.9	91	0.00
23 T	Vinyl Acetate	250.000	247.432	1.0	89	0.00
24 P	1,1-Dichloroethane	50.000	48.687	2.6	90	0.00
25 T	2-Butanone	250.000	262.853	-5.1	96	0.00
26 T	2,2-Dichloropropane	50.000	47.541	4.9	85	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.752	0.5	91	0.00
28 T	Bromochloromethane	50.000	46.281	7.4	84	0.00
29 T	Tetrahydrofuran	250.000	253.987	-1.6	95	0.00
30 C	Chloroform	50.000	49.371	1.3#	92	0.00
31 T	Cyclohexane	50.000	47.487	5.0	90	0.00
32 T	1,1,1-Trichloroethane	50.000	48.373	3.3	87	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.839	4.3	89	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	95	0.00
35 S	Dibromofluoromethane	50.000	50.604	-1.2	91	0.00
36 T	1,1-Dichloropropene	50.000	48.919	2.2	90	0.00
37 T	Ethyl Acetate	50.000	51.840	-3.7	94	0.00
38 T	Carbon Tetrachloride	50.000	49.620	0.8	88	0.00
39 T	Methylcyclohexane	50.000	50.760	-1.5	92	0.00
40 TM	Benzene	50.000	50.738	-1.5	93	0.00
41 T	Methacrylonitrile	50.000	50.582	-1.2	91	0.00
42 TM	1,2-Dichloroethane	50.000	49.261	1.5	90	0.00
43 T	Isopropyl Acetate	50.000	50.828	-1.7	90	0.00
44 TM	Trichloroethene	50.000	51.670	-3.3	93	0.00
45 C	1,2-Dichloropropane	50.000	52.438	-4.9#	94	0.00
46 T	Dibromomethane	50.000	51.344	-2.7	93	0.00
47 T	Bromodichloromethane	50.000	52.182	-4.4	92	0.00
48 T	Methyl methacrylate	50.000	51.826	-3.7	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX012122\
 Data File : VX026537.D
 Acq On : 21 Jan 2022 10:37
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 22 03:56:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	953.681	4.6	85	0.00
50 S	Toluene-d8	50.000	53.049	-6.1	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	265.660	-6.3	96	0.00
52 CM	Toluene	50.000	53.410	-6.8#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	49.532	0.9	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.279	-8.6	93	0.00
55 T	1,1,2-Trichloroethane	50.000	54.431	-8.9	99	0.00
56 T	Ethyl methacrylate	50.000	54.640	-9.3	95	0.00
57 T	1,3-Dichloropropane	50.000	53.926	-7.9	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	264.355	-5.7	95	0.00
59 T	2-Hexanone	250.000	271.439	-8.6	96	0.00
60 T	Dibromochloromethane	50.000	56.326	-12.7	98	0.00
61 T	1,2-Dibromoethane	50.000	54.448	-8.9	98	0.00
62 S	4-Bromofluorobenzene	50.000	55.718	-11.4	101	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	102	0.00
64 T	Tetrachloroethene	50.000	52.315	-4.6	104	0.00
65 PM	Chlorobenzene	50.000	50.709	-1.4	101	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.798	-1.6	99	0.00
67 C	Ethyl Benzene	50.000	51.479	-3.0#	100	0.00
68 T	m/p-Xylenes	100.000	104.437	-4.4	101	0.00
69 T	o-Xylene	50.000	52.646	-5.3	102	0.00
70 T	Styrene	50.000	53.839	-7.7	102	0.00
71 P	Bromoform	50.000	47.899	4.2	101	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	51.508	-3.0	101	0.00
74 T	N-ethyl acetate	50.000	50.281	-0.6	92	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.689	2.6	95	0.00
76 T	1,2,3-Trichloropropane	50.000	50.464	-0.9	97	0.00
77 T	Bromobenzene	50.000	51.353	-2.7	100	0.00
78 T	n-propylbenzene	50.000	52.691	-5.4	102	0.00
79 T	2-Chlorotoluene	50.000	50.824	-1.6	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.525	-5.0	102	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.347	3.3	100	0.00
82 T	4-Chlorotoluene	50.000	51.174	-2.3	100	0.00
83 T	tert-Butylbenzene	50.000	51.825	-3.7	100	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.721	-3.4	100	0.00
85 T	sec-Butylbenzene	50.000	53.434	-6.9	102	0.00
86 T	p-Isopropyltoluene	50.000	53.766	-7.5	102	0.00
87 T	1,3-Dichlorobenzene	50.000	51.040	-2.1	102	0.00
88 T	1,4-Dichlorobenzene	50.000	51.185	-2.4	101	0.00
89 T	n-Butylbenzene	50.000	54.963	-9.9	103	0.00
90 T	Hexachloroethane	50.000	48.789	2.4	100	0.00
91 T	1,2-Dichlorobenzene	50.000	51.232	-2.5	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.669	2.7	90	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.793	-5.6	101	0.00
94 T	Hexachlorobutadiene	50.000	52.887	-5.8	103	0.00
95 T	Naphthalene	50.000	52.344	-4.7	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012122\
Data File : VX026537.D
Acq On : 21 Jan 2022 10:37
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 22 03:56:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.677	-5.4	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6