

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012122\
 Data File : VX026544.D
 Acq On : 21 Jan 2022 13:40
 Operator : JC/MD
 Sample : N1229-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 22 03:58:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	143243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	240566	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94438	47.195	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.400%
35) Dibromofluoromethane	5.391	113	75255	48.442	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.880%
50) Toluene-d8	8.653	98	284933	50.898	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.800%
62) 4-Bromofluorobenzene	11.079	95	101923	51.252	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.500%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012122\
Data File : VX026544.D
Acq On : 21 Jan 2022 13:40
Operator : JC/MD
Sample : N1229-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 22 03:58:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 713
Delta R.T. -0.000 min
Lab File: VX026544.D
Acq: 21 Jan 2022 13:40




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:168 Resp: 143243
Ion Ratio Lower Upper
168 100
99 60.9 50.3 75.5



#33
1,2-Dichloroethane-d4
Concen: 47.195 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.000 min
Lab File: VX026544.D
Acq: 21 Jan 2022 13:40

Tgt Ion: 65 Resp: 94438
Ion Ratio Lower Upper
65 100
67 51.8 0.0 100.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX026544.D
Acq: 21 Jan 2022 13:40




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:114 Resp: 240566
Ion Ratio Lower Upper
114 100
63 21.4 0.0 42.8
88 17.0 0.0 33.0



#35
Dibromofluoromethane
Concen: 48.442 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX026544.D
Acq: 21 Jan 2022 13:40

Tgt Ion:113 Resp: 75255
Ion Ratio Lower Upper
113 100
111 102.4 81.2 121.8
192 19.1 15.0 22.4









#50
Toluene-d8
Concen: 50.898 ug/l
RT: 8.653 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026544.D
Acq: 21 Jan 2022 13:40




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion: 98 Resp: 284933
Ion Ratio Lower Upper
98 100
100 65.7 53.2 79.8



#62
4-Bromofluorobenzene
Concen: 51.252 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX026544.D
Acq: 21 Jan 2022 13:40

Tgt Ion: 95 Resp: 101923
Ion Ratio Lower Upper
95 100
174 73.8 0.0 148.6
176 71.7 0.0 141.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX026544.D
Acq: 21 Jan 2022 13:40

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Tgt Ion:117 Resp: 216925
Ion Ratio Lower Upper
117 100
82 57.0 45.2 67.8
119 31.9 26.2 39.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX026544.D
Acq: 21 Jan 2022 13:40

Tgt Ion:152 Resp: 89031
Ion Ratio Lower Upper
152 100
115 61.4 44.2 132.6
150 156.0 0.0 348.2

