

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012125\
 Data File : VX044680.D
 Acq On : 21 Jan 2025 12:46
 Operator : JC/MD
 Sample : VIBLK689
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK689

Quant Time: Jan 22 01:37:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012125WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 22 01:17:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.751	114	298501	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	258391	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	106651	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	126406	49.068	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.140%
7) Chloroethane-d5	1.648	69	29849	44.975	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.960%
11) 1,1-Dichloroethene-d2	2.294	65	52932	54.423	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	108.840%
21) 2-Butanone-d5	4.452	46	132352	102.488	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.490%
24) Chloroform-d	5.050	84	215946	50.717	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.440%
26) 1,2-Dichloroethane-d4	5.946	65	139613	49.715	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.420%
32) Benzene-d6	5.964	84	449805	52.161	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.320%
36) 1,2-Dichloropropane-d6	7.299	67	143559	52.670	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.340%
41) Toluene-d8	8.647	98	398120	51.643	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.280%
43) trans-1,3-Dichloroprop...	8.945	79	71750	51.363	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.720%
47) 2-Hexanone-d5	9.384	63	85498	109.184	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.180%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	178389	54.294	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.580%
66) 1,2-Dichlorobenzene-d4	12.317	152	123306	53.801	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012125\
Data File : VX044680.D
Acq On : 21 Jan 2025 12:46
Operator : JC/MD
Sample : VIBLK689
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK689

Quant Time: Jan 22 01:37:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012125WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 22 01:17:07 2025
Response via : Initial Calibration

