

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012125\
 Data File : VX044686.D
 Acq On : 21 Jan 2025 17:18
 Operator : JC/MD
 Sample : PB166134ZHE#11 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB166134ZHE#11

Quant Time: Jan 22 01:39:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012125WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 22 01:17:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	292693	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	250535	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	100901	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	129650	51.326	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.660%
7) Chloroethane-d5	1.648	69	28116	43.205	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.400%
11) 1,1-Dichloroethene-d2	2.294	65	43126	45.220	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.440%
21) 2-Butanone-d5	4.452	46	136784	108.022	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.020%
24) Chloroform-d	5.050	84	220999	52.934	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.860%
26) 1,2-Dichloroethane-d4	5.946	65	142340	51.691	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.380%
32) Benzene-d6	5.964	84	461468	55.192	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.380%
36) 1,2-Dichloropropane-d6	7.300	67	148953	56.362	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	112.720%
41) Toluene-d8	8.641	98	395130	52.863	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.720%
43) trans-1,3-Dichloroprop...	8.946	79	72864	53.797	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	107.600%
47) 2-Hexanone-d5	9.384	63	91167	120.074	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	120.070%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	185185	58.129	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	116.260%
66) 1,2-Dichlorobenzene-d4	12.317	152	120756	55.691	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012125\
Data File : VX044686.D
Acq On : 21 Jan 2025 17:18
Operator : JC/MD
Sample : PB166134ZHE#11 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB166134ZHE#11

Quant Time: Jan 22 01:39:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM012125WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 22 01:17:07 2025
Response via : Initial Calibration

