

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012125\
 Data File : VX044685.D
 Acq On : 21 Jan 2025 16:55
 Operator : JC/MD
 Sample : PB166134TB 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VLEB134

Quant Time: Jan 22 01:38:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012125WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 22 01:17:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	307961	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	268820	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	107192	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	125384	47.176	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	94.360%	
7) Chloroethane-d5	1.648	69	25009	36.525	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	73.060%	
11) 1,1-Dichloroethene-d2	2.294	65	49252	49.083	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	98.160%	
21) 2-Butanone-d5	4.458	46	134617	101.040	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	101.040%	
24) Chloroform-d	5.050	84	215164	48.981	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.960%	
26) 1,2-Dichloroethane-d4	5.946	65	139292	48.077	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.160%	
32) Benzene-d6	5.964	84	441931	49.260	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.520%	
36) 1,2-Dichloropropane-d6	7.299	67	144443	50.938	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	101.880%	
41) Toluene-d8	8.647	98	389274	48.537	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	97.080%	
43) trans-1,3-Dichloroprop...	8.945	79	72255	49.718	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	99.440%	
47) 2-Hexanone-d5	9.384	63	86966	106.750	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	106.750%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	178986	52.362	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	104.720%	
66) 1,2-Dichlorobenzene-d4	12.317	152	119706	51.966	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	103.940%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012125\
Data File : VX044685.D
Acq On : 21 Jan 2025 16:55
Operator : JC/MD
Sample : PB166134TB 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VLEB134

Quant Time: Jan 22 01:38:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012125WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 22 01:17:07 2025
Response via : Initial Calibration

