

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012125\
 Data File : VX044687.D
 Acq On : 21 Jan 2025 17:41
 Operator : JC/MD
 Sample : Q1130-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JNLA7

Quant Time: Jan 22 01:39:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML012125WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 22 01:17:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.751	114	379339	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	326475	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	135434	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	200796	61.334	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 122.660%	
7) Chloroethane-d5	1.648	69	43783	51.912	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 103.820%	
11) 1,1-Dichloroethene-d2	2.300	65	64814	52.438	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 104.880%	
21) 2-Butanone-d5	4.452	46	216454	131.895	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 131.890%#	
24) Chloroform-d	5.050	84	347321	64.189	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 128.380%#	
26) 1,2-Dichloroethane-d4	5.946	65	222234	62.271	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 124.540%	
32) Benzene-d6	5.964	84	717010	65.807	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 131.620%#	
36) 1,2-Dichloropropane-d6	7.299	67	233048	67.671	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 135.340%#	
41) Toluene-d8	8.641	98	621267	63.783	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 127.560%#	
43) trans-1,3-Dichloroprop...	8.945	79	114878	65.087	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 130.180%#	
47) 2-Hexanone-d5	9.384	63	139265	140.758	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 140.760%#	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	285701	68.821	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 137.640%#	
66) 1,2-Dichlorobenzene-d4	12.317	152	196435	67.493	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 134.980%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012125\
Data File : VX044687.D
Acq On : 21 Jan 2025 17:41
Operator : JC/MD
Sample : Q1130-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JNLA7

Quant Time: Jan 22 01:39:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM012125WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 22 01:17:07 2025
Response via : Initial Calibration

