

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012219\
 Data File : VX007183.D
 Acq On : 22 Jan 2019 15:55
 Operator : JC/SP
 Sample : K1214-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW526-4246

Quant Time: Jan 23 01:20:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	497809	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	771733	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	725534	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	363760	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	269243	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
35) Dibromofluoromethane	5.50	113	241786	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
50) Toluene-d8	8.71	98	917048	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	11.13	95	333850	52.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.88%	
Target Compounds						
12) 1,1-Dichloroethene	2.38	96	4262	0.932	ug/l #	88
24) 1,1-Dichloroethane	3.70	63	280305	34.006	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012219\
Data File : VX007183.D
Acq On : 22 Jan 2019 15:55
Operator : JC/SP
Sample : K1214-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW526-4246

Quant Time: Jan 23 01:20:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

