

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012219\
 Data File : VX007188.D
 Acq On : 22 Jan 2019 18:08
 Operator : JC/SP
 Sample : K1214-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW531-2327

Quant Time: Jan 23 01:32:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	514300	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	797917	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	756846	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	371080	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	277590	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
35) Dibromofluoromethane	5.51	113	248234	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
50) Toluene-d8	8.71	98	967013	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
62) 4-Bromofluorobenzene	11.13	95	341583	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
Target Compounds						
16) Acetone	2.45	43	6366	2.396	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012219\
Data File : VX007188.D
Acq On : 22 Jan 2019 18:08
Operator : JC/SP
Sample : K1214-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW531-2327

Quant Time: Jan 23 01:32:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

