

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012219\
 Data File : VX007189.D
 Acq On : 22 Jan 2019 18:35
 Operator : JC/SP
 Sample : K1214-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW531-3337

Quant Time: Jan 23 01:34:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	498043	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	778528	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	729124	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	358717	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	272112	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
35) Dibromofluoromethane	5.51	113	242482	48.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.28%	
50) Toluene-d8	8.71	98	929647	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
62) 4-Bromofluorobenzene	11.13	95	328009	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
Target Compounds						
16) Acetone	2.45	43	7977	3.100	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012219\
Data File : VX007189.D
Acq On : 22 Jan 2019 18:35
Operator : JC/SP
Sample : K1214-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW531-3337

Quant Time: Jan 23 01:34:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

