

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012219\
 Data File : VX007190.D
 Acq On : 22 Jan 2019 19:01
 Operator : JC/SP
 Sample : K1214-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW531-4246

Quant Time: Jan 28 03:36:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	503846	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	780050	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	742436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	366889	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	277242	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
35) Dibromofluoromethane	5.51	113	241044	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
50) Toluene-d8	8.71	98	940979	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	11.13	95	337379	52.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.86%	
Target Compounds						
16) Acetone	2.45	43	15687	6.026	ug/l #	88
52) Toluene	8.79	92	15896	1.195	ug/l	96
69) o-Xylene	10.70	106	4824	0.470	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012219\
Data File : VX007190.D
Acq On : 22 Jan 2019 19:01
Operator : JC/SP
Sample : K1214-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW531-4246

Quant Time: Jan 28 03:36:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

